

Guide for Distributors of Medical Devices



CONTENTS

1	SCOPE	3
2	INTRODUCTION	3
3	LEGISLATIVE BASIS	4
4	TRACEABILITY AND UNIQUE DEVICE IDENTIFICATION (UDI) SYSTEM	10
4.1	Documentation and record keeping	10
4.2	Unique Device Identification (UDI) System	11
4.3	Falsified medical devices	11
5	EUDAMED AND REGISTRATION	12
6	MEDICAL DEVICE RECEIPT, STORAGE, SUPPLY AND TRANSPORT	13
6.1	Introduction of a new medical device	13
6.2	Placing devices into saleable stock	13
6.3	Storage	15
6.4	Pest control	16
6.5	Medical device disposal	16
6.6	Transport	17
7	SYSTEM AND PROCEDURE PACKS	17
8	PARTS AND COMPONENTS	18
9	FIELD SAFETY CORRECTIVE ACTIONS (FSCA)	19
9.1	Medical device recalls	20
10	COMPLAINTS	21
11	THE MANAGEMENT OF RETURNED MEDICAL DEVICES	21
12	PERSONNEL AND STAFF TRAINING	23
13	IMPLEMENTATION OF A QUALITY SYSTEM	23
13.1	Change control	25
13.2	Quality risk management	25
13.3	Non-conformances, investigations and corrective and preventive actions (CAPA)	26
13.4	Validation and qualification	27
13.5	Outsourced activities	28
13.6	Internal audits	28
14	FURTHER INFORMATION	29
	APPENDIX 1 FURTHER GUIDANCE	30

1 SCOPE

The purpose of this document is to provide guidance to industry on the distribution of medical devices, *in vitro* diagnostic medical devices, and their accessories for human use in Ireland. For the purpose of this document, all such medical devices are referred to collectively as 'medical devices' unless otherwise specified. Additional obligations apply to distributors who also function as importers into the European Union. Further information regarding the obligations of importers of medical devices can be found on the Health Products Regulatory Authority (HPRA) website www.hpra.ie.

2 INTRODUCTION

The safety and performance of medical devices directly affects patient safety and health. The range of medical devices available is vast and they are used by a broad variety of users from the general public in their homes to the most critically ill patients in specialist clinical settings. Patients using devices include extremely vulnerable groups such as neonates and infants, the elderly, people with disabilities and other patients who are particularly susceptible to diseases. Medical devices may also be used in high-risk surgical procedures and intensive care settings, where improper storage along the supply chain, amongst other aspects, may lead to undesirable, and in some cases extremely serious, consequences.

[Regulation 2017/745](#) on medical devices and [Regulation 2017/746](#) on *in vitro* diagnostic devices (together known as 'the Regulations') entered into force on 25 May 2017. The Regulations have a staggered transitional period. The obligations for distributors will apply with the full application of the Medical Device Regulation (MDR) in May 2021 and full application of the Regulation on *in vitro* diagnostic medical devices (IVDR) in May 2022.

These two Regulations replace the European Directives for medical devices, including Council Directives 90/385/EEC, 93/42/EEC and 98/79/EC. The requirements relating to distribution of devices are essentially the same in both the MDR and IVDR.

This document aims to specify the obligations for distributors as outlined in the Regulations, along with supplementary guidance and frameworks that can be used by distributors to meet those obligations. In addition, it sets out the HPRA's recommendations for good distribution practices. A distributor may have alternative methods for achieving compliance with the regulations and, if justified, these alternatives may be equally acceptable.

3 LEGISLATIVE BASIS

At both national and European level, the legislative basis for distribution of medical devices is based on the Regulations. The definitions relevant to medical device distributors are identified in table 1 below.

Table 1: Definitions as per the Regulations

TERM	DEFINITION
Authorised representative	any natural or legal person established within the Union who has received and accepted a written mandate from a manufacturer, located outside the Union, to act on the manufacturer's behalf in relation to specified tasks with regard to the latter's obligations under this Regulation
Distributor	any natural or legal person in the supply chain, other than the manufacturer or the importer, that makes a device available on the market up until the point of putting into service Note: Recitals 28 of the MDR and 26 of the IVDR also state that for the purpose of these Regulations, the activities of distributors should be deemed to include acquisition, holding and supplying of devices.
Field safety corrective action	corrective action taken by a manufacturer for technical or medical reasons to prevent or reduce the risk of a serious incident in relation to a device made available on the market
Field safety notice	a communication sent by a manufacturer to users or customers in relation to a field safety corrective action
HPRA safety notice	a communication sent by the HPRA to advise the user of the device, whether that is at home, in a hospital or in a community setting, of important information regarding the safe use of their medical device
Importer	any natural or legal person established within the Union that places a device from a third country on the Union market

TERM	DEFINITION
Making available on the market	any supply of a device, other than an investigational device, for distribution, consumption or use on the Union market in the course of a commercial activity, whether in return for payment or free of charge
Manufacturer	a natural or legal person who manufactures or fully refurbishes a device or has a device designed, manufactured or fully refurbished, and markets that device under its name or trademark
Placing on the market	the first making available of a device, other than an investigational device, on the Union market
Putting into service	the stage at which a device, other than an investigational device, has been made available to the final user as being ready for use on the Union market for the first time for its intended purpose
Recall	any measure aimed at achieving the return of a device that has already been made available to the end user
Withdrawal	any measure aimed at preventing a device in the supply chain from being further made available on the market

Article 14 of the Regulations describes the general obligations of medical device distributors. Article 16 of the Regulations details cases in which obligations of manufacturers apply to importers, distributors or other persons as outlined below. Article 25 of the MDR and Article 22 of the IVDR describe the obligations relating to identification and traceability within the supply chain. The obligations in these articles relating to distributors are included in table 2 below.

Table 2: Distributors’ obligations
(Further details of the Articles referenced in this table can be found in Appendix 1)

MDR AND IVDR ARTICLE 14	OBLIGATION
Paragraph 1	When making a device available on the market, distributors shall, in the context of their activities, act with due care in relation to the requirements applicable.
Paragraph 2	Before making a device available on the market, distributors shall verify that all of the following requirements are met:

MDR AND IVDR ARTICLE 14	OBLIGATION
	<p>a) the device has been CE marked and that the EU declaration of conformity of the device has been drawn up;</p> <p>b) the device is accompanied by the information to be supplied by the manufacturer in accordance with Article 10(11) [<i>Article 10(10) of the IVD Regulation</i>];</p> <p>c) for imported devices, the importer has complied with the requirements set out in Article 13(3);</p> <p>d) that, where applicable, a UDI* has been assigned by the manufacturer.</p> <p><i>*A UDI is a unique device identifier.</i></p>
	<p>In order to meet the requirements referred to in points (a), and (b) and (d) of the first subparagraph the distributor may apply a sampling method representative of devices supplied by that distributor.</p>
	<p>Where a distributor considers or has reason to believe that a device is not in conformity with the requirements of this Regulation, it shall not make the device available on the market until it has been brought into conformity, and shall inform the manufacturer and, where applicable, the manufacturer's authorised representative, and the importer. Where the distributor considers or has reason to believe that the device presents a serious risk or is a falsified device, it shall also inform the competent authority of the Member State in which it is established.</p>
<p>Paragraph 3</p>	<p>Distributors shall ensure that, while the device is under their responsibility, storage or transport conditions comply with the conditions set by the manufacturer.</p>
<p>Paragraph 4</p>	<p>Distributors that consider or have reason to believe that a device which they have made available on the market is not in conformity with this Regulation shall immediately inform the manufacturer and, where applicable, the manufacturer's authorised representative and the importer. Distributors shall co-operate with the manufacturer and, where applicable, the manufacturer's authorised representative, and the importer, and with competent authorities to ensure that the necessary corrective action to bring that device into conformity, to withdraw or to recall it, as appropriate, is taken. Where the distributor considers or has reason to believe that the device presents a serious risk, it shall also immediately inform the competent authorities of the Member States in which it made the device available, giving details, in particular, of the non-compliance and of any corrective action taken.</p>

MDR AND IVDR ARTICLE 14	OBLIGATION
Paragraph 5	<p>Distributors that have received complaints or reports from healthcare professionals, patients or users about suspected incidents related to a device they have made available, shall immediately forward this information to the manufacturer and, where applicable, the manufacturer's authorised representative, and the importer. They shall keep a register of complaints, of non-conforming devices and of recalls and withdrawals, and keep the manufacturer and, where available, the authorised representative and the importer informed of such monitoring and provide them with any information upon their request.</p>
Paragraph 6	<p>Distributors shall, upon request by a competent authority, provide it with all the information and documentation that is at their disposal and is necessary to demonstrate the conformity of a device.</p> <p>Distributors shall be considered to have fulfilled the obligation referred to in the first subparagraph when the manufacturer or, where applicable, the authorised representative for the device in question provides the required information. Distributors shall co-operate with competent authorities, at their request, on any action taken to eliminate the risks posed by devices which they have made available on the market. Distributors, upon request by a competent authority, shall provide free samples of the device or, where that is impracticable, grant access to the device.</p>
MDR AND IVDR ARTICLE 16	OBLIGATION
Paragraph 1	<p>A distributor, importer or other natural or legal person shall assume the obligations incumbent on manufacturers if it does any of the following:</p> <ul style="list-style-type: none"> a) makes available on the market a device under its name, registered trade name or registered trade mark, except in cases where a distributor or importer enters into an agreement with a manufacturer whereby the manufacturer is identified as such on the label and is responsible for meeting the requirements placed on manufacturers in this Regulation; b) changes the intended purpose of a device already placed on the market or put into service; c) modifies a device already placed on the market or put into service in such a way that compliance with the applicable requirements may be affected.

MDR AND IVDR ARTICLE 16	OBLIGATION
	<p>The first subparagraph shall not apply to any person who, while not considered a manufacturer as defined in point (30) of Article 2, assembles or adapts for an individual patient a device already on the market without changing its intended purpose.</p>
<p>Paragraph 2</p>	<p>For the purposes of point (c) in Article 16(1), the following shall not be considered to be a modification of a device that could affect its compliance with the applicable requirements:</p> <ul style="list-style-type: none"> a) provision, including translation, of the information supplied by the manufacturer in accordance with section 23 of Annex I, relating to a device already placed on the market and of further information which is necessary in order to market the product in the relevant Member State; b) changes to the outer packaging of a device already placed on the market, including a change of pack size, if the repackaging is necessary in order to market the device in the relevant Member State and if it is carried out in such conditions that the original condition of the device cannot be affected by it. In the case of devices placed on the market in sterile condition, it shall be presumed that the original condition of the device is adversely affected if the packaging that is necessary for maintaining the sterile condition is opened, damaged or otherwise negatively affected by the repackaging
<p>Paragraph 3</p>	<p>A distributor or importer that carries out any of the activities mentioned in points (a) and (b) of paragraph 2 shall indicate on the device or, where that is impracticable, on its packaging or in a document accompanying the device, the activity carried out together with its name, registered trade name or registered trade mark, registered place of business and the address at which it can be contacted, so that its location can be established.</p> <p>Distributors and importers shall ensure that they have in place a quality management system that includes procedures which ensure that the translation of information is accurate and up-to-date, and that the activities mentioned in points (a) and (b) of paragraph 2 are performed by a means and under conditions that preserve the original condition of the device and that the packaging of the repackaged device is not defective, of poor quality or untidy. The quality management system shall cover, <i>inter alia</i>, procedures ensuring that the distributor or importer is informed of any corrective action taken by the manufacturer in relation to the device in question in order to respond to safety issues or to bring it into conformity with this Regulation.</p>

MDR AND IVDR ARTICLE 16	OBLIGATION
Paragraph 4	<p>At least 28 days prior to making the relabelled or repackaged device available on the market, distributors or importers carrying out any of the activities mentioned in points (a) and (b) of paragraph 16 (2) shall inform the manufacturer and the competent authority of the Member State in which they plan to make the device available of the intention to make the relabelled or repackaged device available and, upon request, shall provide the manufacturer and the competent authority with a sample or mock-up of the relabelled or repackaged device, including any translated label and instructions for use.</p> <p>Within the same period of 28 days, the distributor or importer shall submit to the competent authority a certificate, issued by a notified body designated for the type of devices that are subject to activities mentioned in points (a) and (b) of paragraph 16 (2), attesting that the quality management system of the distributor or importer complies with the requirements laid down in paragraph 16 (3).</p>

MDR AND IVDR ARTICLE 22	OBLIGATION
Paragraph 1	<p>Distributors and importers shall co-operate with manufacturers or authorised representatives to achieve an appropriate level of traceability of devices.</p>
Paragraph 2	<p>Economic operators shall be able to identify the following to the competent authority, for the period referred to in Article 10(8)* [<i>Article 10(7) of the IVD Regulation</i>]:</p> <ul style="list-style-type: none"> a) any economic operator to whom they have directly supplied a device; b) any economic operator who has directly supplied them with a device; c) any health institution or healthcare professional to which they have directly supplied a device. <p><i>* Records must be kept for period of at least ten years after the last device covered by the EU declaration of conformity has been placed on the market. In the case of implantable devices, the period shall be at least 15 years after the last device has been placed on the market.</i></p>

4 TRACEABILITY AND UNIQUE DEVICE IDENTIFICATION (UDI) SYSTEM

Article 25 of the MDR and Article 22 of the IVDR relate to the obligations of distributors to achieve an appropriate level of traceability of medical devices that they distribute.

Medical device traceability is achieved through maintaining adequately detailed records in relation to the sourcing and supply of medical devices. For example, in the event of a Field Safety Corrective Action (FSCA) it may be necessary to determine the customers that received the medical device that was affected by the FSCA. In such cases, the maintenance of a system which includes tracking by batch/lot number is most valuable in terms of assisting and ensuring the swift conduct of the FSCA and limiting it to the affected batch/lot only.

Other systems for traceability and recall are available and may be justified, e.g. tracking by date. The system used to maintain product traceability should be challenged periodically to ensure that it is capable of determining stock location.

4.1 Documentation and record keeping

Distributors should keep adequate records, including records of customers and suppliers to and from whom medical devices have been distributed. Records should include but are not limited to: medical device name and/or code, batch or lot number, quantities and delivery date. This information is especially important in the event of an FSCA. In the event that the FSCA requires the recall or withdrawal of a particular batch/lot, failure to record the batch/lot numbers involved in transactions may result in a full product recall.

Distributors must co-operate with manufacturers or authorised representatives to achieve an appropriate level of traceability of devices. They should be able to identify the following to the competent authority, for the retention period applicable:

- to whom they have directly supplied a device, including healthcare professionals, institutions and sales representatives
- who has directly supplied them with a device

Records to be maintained by the distributor could include:

- copies of invoices relating to the receipt and supply of a medical device
- copies of orders relating to the receipt and supply of a medical device
- a list of approved medical device suppliers and details of the relevant medical devices
- a list of approved customers, for example economic operators, health institutions and healthcare professionals, including contact details of those to whom medical devices were supplied
- records of checks carried (for example, labelling checks for CE marks) and the approval of medical devices into saleable stock

These records should be kept for ten years after the last device covered by the EU declaration of conformity has been placed on the market. In the case of implantable devices, the period must

be at least 15 years after the last device has been placed on the market. These records may be requested by the HPRA.

4.2 Unique Device Identification (UDI) System

Article 27(8) of the MDR and Article 24(8) of the IVDR require that distributors store and keep, preferably by electronic means, the UDI of the devices which they have supplied or with which they have been supplied, if those devices are class III implantable devices or belong to one of the categories of devices specified by the European Commission as requiring this traceability.

Further information and guidance on application of UDI is available on the European Commission [website](#).

4.3 Falsified medical devices

It is imperative that all distributors operate using good governance and are vigilant in their efforts to prevent falsified medical devices from being traded with other distributors or placed on the market. It is recommended that distributors put a process and related procedure in place to be followed in the event of identifying a suspected falsified medical device or of being notified that a (suspected) falsified medical device has been received. Training should be arranged for staff on following this process and procedure.

Distributors should consider the following to help prevent the introduction of falsified medical devices into the supply chain:

- Be aware of the possibility of falsified medical devices being supplied inadvertently through legitimate sources.
- Have robust systems for ensuring the legitimacy of suppliers and ensure this is regularly reviewed.
- Maintain a list of approved suppliers and ensure medical devices are only sourced directly from these approved suppliers. In this regard the approval process should include assessment of supply chain and where possible the authority of the supplier to supply the specific medical devices, e.g. if devices are not received directly from the manufacturer, seek confirmation from the manufacturer or official distribution partners.
- Be familiar with the history of the supply chain for medical devices received and question previous stages in the supply chain, if deemed necessary.
- Train staff to be aware of falsified medical devices and what to look out for.
- Ensure that the goods-in procedure involves a detailed inspection of medical devices received, which is capable of identifying changes or unusual aspects to the appearance and packaging of medical devices.
- Be familiar with the market price of the medical devices they source and normal fluctuations in this price. Offers below expected fluctuation should be treated cautiously and investigated to ensure these are genuine.
- Be vigilant and do not allow themselves to be used by counterfeiters to 'launder' falsified medical devices.

- Be aware of the possibility of falsified medical device entering the supply chain through returns.
- Be knowledgeable of medical devices at risk of counterfeiting and ensure purchasers for the distributor are aware of such medical devices also.

A distributor in possession of a medical device that is found to be (or is suspected of being) falsified is responsible for the withdrawal and quarantine of the medical device from saleable stock. If a distributor located in Ireland is suspicious that a medical device that is being offered or that has been received is not genuine, then the distributor should inform the HPRA immediately. Medical devices which are suspected of being falsified should never be returned to the supplier without the consent of the HPRA.

Any suspicious approaches or activities noticed by a distributor should be reported to the HPRA without delay (devices@hpra.ie).

5 EUDAMED AND REGISTRATION

Eudamed is the IT system developed by the European Commission to ensure effective implementation of the MDR and IVDR.

Eudamed will include six interconnected modules:

- Actor registration, including a single registration number (SRN)
- UDI/device registration
- Notified bodies and certificates
- Clinical investigations and performance studies
- Vigilance and post-market surveillance
- Market surveillance

It is envisaged that all of the modules of Eudamed and the audit of the system will be completed before May 2022.

The Commission deployed the actor registration module of Eudamed on 01 December 2020. Actors that register on Eudamed will be able to obtain a SRN. Further details regarding the actor registration module can be found on the EU Commission [website](#).

While distributors are not required to register themselves as actors on Eudamed, they will be able to use Eudamed to carry out some of the essential verification checks required under Article 14 once Eudamed is fully functional. Distributors that qualify as importers (e.g. when importing medical devices from outside of the EU) will be required to register as an actor on Eudamed. Distributors established in Ireland are encouraged to notify the HPRA of their establishment through our economic operator portal available on our [website](#). Information on national registration requirements will be updated on our Registration [webpage](#).

6 MEDICAL DEVICE RECEIPT, STORAGE, SUPPLY AND TRANSPORT

6.1 Introduction of a new medical device

A robust system for approval of new medical device suppliers is a key component in the prevention of falsified medical devices entering the supply chain. It is recommended that distributors have a procedure in place relating to the introduction of new medical devices to their inventory. The required storage and transport conditions and any additional requirements relating to their distribution should be known and documented. Where clarification is required, distributors should consult with their suppliers with respect to the classification of new medical devices.

Documentation relating to the introduction of each new medical device should be available for review. Specific documentation required for placing a device into saleable stock is detailed in section 6.2 of this document.

Distributors receiving medical devices directly from operators in a country outside of the European Economic Area may be considered to be importers. In accordance with legislation, importers have specific legal obligations in addition to a distributor's obligations.

The Regulations also state that where a distributor performs certain activities, they may assume the responsibility of a manufacturer (see section 3 and Appendix 1 for further details). These activities include where a distributor:

- makes a device available on the market under their name, registered trade name or registered trade mark
- changes the intended purpose of the device already placed on the market or put into service
- modifies a device already placed on the market or put into service in such a way as to affect the compliance of the device

6.2 Placing devices into saleable stock

It is recommended that deliveries are examined at receipt to include checking for damage (including any breaches to sterile packaging) and appropriate remaining shelf life. The system in place to enable detection and quarantine of any non-conforming or defective product should be documented.

There should be a system in place to enable the distributor to meet the requirements of Article 14 paragraph 2 of the Regulations (see table 2). Before making the medical device available on the market, the distributor must verify that certain minimum requirements are met. This includes checking that a CE mark is present and that a declaration of conformity of the device has been drawn up. In performing this verification, the distributor may apply an appropriate sampling method that ensures that the samples selected are representative of the devices supplied by that distributor. Any sampling method applied should be risk-based and justified. Expiry dates should also be checked to ensure that expired medical devices are not supplied. For increased

traceability it is recommended that batch or lot numbers are recorded. All the above mentioned checks should be recorded.

In addition, distributors must also verify that the following requirements are met:

- the medical device is accompanied by the information to be supplied by the manufacturer in accordance with Article 10(11) of the MDR and Article 10(10) of the IVDR (i.e. labelling and instructions for use). A risk-based sampling approach can be used for this check;
- for imported devices, the importer has complied with the requirements set out in Article 13(3) of the Regulations (i.e. importer's contact details);
- where applicable, a unique device identifier (UDI) has been assigned by the manufacturer.

The verification checks should also be recorded.

The process for handling of non-conforming goods at goods-in should be documented, including where the non-conforming medical device is stored, what documentation is completed and how the stock is controlled on the warehouse management system (if applicable).

It should be verified that all goods have been received from an approved supplier. To ensure this, goods-in personnel should have access to the list of suppliers approved under the company's quality system. Alternatively, an inventory management system may be used which permits medical devices to be booked onto the inventory system only if the supplier is approved to supply that medical device.

Specific checks should be performed on medical devices requiring refrigerated storage. These checks should include, but are not limited to: checking that cold chain conditions were maintained during transportation, and checking that the consignment(s) were received within the validated transportation time if received in qualified cold-chain shippers.

If a medical device is received under quarantine status, there should be systems and procedures in place to ensure that it is not released into saleable stock until all necessary conditions have been met.

Where a distributor considers or has reason to believe that a device is not in conformity with the requirements of the Regulations, they must ensure that the device is brought into conformity before it becomes available on the market. It is also a requirement that in this case, the distributor informs the manufacturer and, where applicable, the authorised representative and the importer.

Where the distributor considers or has reason to believe that the medical device presents a serious risk to patient health and safety or is falsified, they shall also inform the HPR. If the distributor has made a medical device that poses a serious risk available in other markets within the EEA, the distributor is also required to inform the national competent authority in each of those markets. This requires that the distributor has a system in place relating to identification of

serious risk and of falsified medical devices. Further information regarding the vigilance system for medical devices is available on the HPRA [website](#).

Records should be available of all checks performed and these should be available to the HPRA during an inspection.

6.3 Storage

According to Article 14(3) of the Regulations, distributors must ensure that, while medical devices are under their responsibility, storage or transport conditions comply with the conditions set by the manufacturer. Storage conditions are normally specified on the containers, for example 'Store between 2 and 8°C', 'Store at -20°C' or 'Do not store above 25°C'. Medical devices must be stored in accordance with the labelled conditions, including potential relative humidity conditions. Where there are no specified storage conditions, distributors should verify with the manufacturer that there are no restrictions on the temperatures at which medical devices may be stored.

Continuous temperature monitoring must be performed and documented in order to ensure that the appropriate conditions are maintained, if applicable. This applies to all areas where medical devices are stored (e.g. bulk storage, pick-face, quarantine and returns areas). At a minimum, a calibrated max/min type thermometer should be used. The maximum and minimum temperatures reached should be recorded every day and the thermometer reset after the readings have been taken.

Temperature monitoring records should be reviewed and approved regularly to ensure compliance with the required conditions. Temperature excursions should be investigated immediately and documented. Consideration should be given to the potential impact that the excursion may have had on devices stored during that time period. In the event that there are multiple temperature excursions across a time period, the impact of the multiple excursions on the devices should be considered. The manufacturer of the medical device should be consulted to ascertain the effect of any excursions from the labelled storage conditions. The method for investigating excursions should be documented.

The distributor should review calibration certificates for temperature monitoring devices to ensure that the accuracy of the devices is acceptable. Documentation should be available for inspection demonstrating that this review has occurred.

Temperatures should also be monitored during periods when temperature measuring devices used routinely are being recalibrated. This may mean that auxiliary temperature measuring devices are required. These devices should also be calibrated.

Adequate rotation of stock (a 'first expired first out' (FEFO) or equivalent stock rotation system) should be applied to ensure that stock with adequate expiry is supplied.

Quarantined stock should be kept separate from approved stock. There should be a system in place to ensure that quarantined stock is not available for picking or returned to saleable stock inadvertently. An inventory of quarantined stock may be maintained on a warehouse management system. The distributor should exercise care to ensure that, where segregation is achieved only by using a computerised system, this system is at least as effective as if devices were physically segregated. This is to ensure that devices that have been deemed as not suitable for saleable stock (e.g. falsified, recalled or defective medical devices) are not supplied to customers.

Any samples received by sales representatives should be stored and transported in accordance with the labelled conditions.

6.4 Pest control

The distributor is responsible for preserving the device through correct storage and transportation as applicable. Therefore, it is recommended that a pest control programme is in place. At a minimum this should include rodent control. Further controls may be required for birds and flying and crawling insects and may include electrical fly killers, glue traps, etc. Rodent control should cover both internal and external locations.

The pest control programme should be documented. A bait map should show the locations of all pest control monitoring stations and should be approved by the distributor.

Any recommendations made by a pest control service provider should be completed and recorded. If recommendations are not completed, an explanation should be recorded. All pest control records should be approved by the distributor and maintained.

6.5 Medical device disposal

Disposal of medical devices should be controlled such that they cannot re-enter the supply chain. All medical devices that are rejected in-house, rejected when received as a return from a customer or recalled stock should, if instructed accordingly, be destroyed in an appropriate and timely manner and in accordance with waste legislation and any associated manufacturer instructions as provided in the product labelling and instructions for use. The manufacturer should be consulted if required. The decision to dispose of medical devices should be documented and recorded.

There should be an inventory of medical devices placed into waste. Records and certificates of destruction should be maintained.

Service level agreements should be in place with third party contractors.

6.6 Transport

Medical devices must be transported appropriately and in accordance with labelled storage conditions (including sterile conditions if applicable) and written procedures. Where contract service providers are used, the distributor must make itself aware of the operating procedures of that party (e.g. by audit). This assessment should include examination of the transportation methods and routes. The distributor should be fully aware of, and agree to, any operations sub-contracted to another party by the contract service provider. The contracted arrangements for transportation should be documented in a service level agreement, and should include details of any sub-contracting.

The distributor must ensure that medical devices are not subjected to prolonged periods of storage during transportation.

7 SYSTEM AND PROCEDURE PACKS

Distributors should be familiar with Article 22 of the MDR which outlines the obligations of a system and procedure pack producer (SPP producer). System and Procedure Packs (SPP) are defined in Article 2 of MDR 2017/745:

'procedure pack' means a combination of products packaged together and placed on the market with the purpose of being used for a specific medical purpose;

'system' means a combination of products, either packaged together or not, which are intended to be interconnected or combined to achieve a specific medical purpose;

For the purpose of this guidance, the term 'system and procedure pack producer' refers to the natural or legal person referred to in Article 22(1), 22(2) and 22(3) of the Medical Device Regulation.

Distributors will be consider a SPP producer when combining a medical device with:

- other devices bearing a CE mark;
- *in vitro* diagnostic medical devices (IVDs) bearing a CE mark in conformity with 2017/746;
- other products which are in conformity with legislation that applies to those products only, (e.g. medicines, personal protective equipment, biocides, etc.) where they are used within a medical procedure or where their presence within the system or procedure pack is justified

Under these circumstances the SPP producer is not required to affix an additional CE mark to the system or procedure pack. However, SPP producers are required to affix their details to the pack including name, registered trade name or registered trademark (alongside contact details). This information should **not** replace or hide the details of the legal manufacturer. System and procedure packs should also be accompanied by the information outlined in Section 23 of Annex I (Information to be Supplied With a Device).

Article 22 of the MDR calls out the need for a SPP producer to draw up a statement if carrying out these activities. This statement needs to be kept available for the national competent authority (i.e. the HPRA) for a period of 10 or 15 years (as applicable under Article 10(8)).

In order to facilitate traceability, EU Commission guidance has been developed to address registration requirements and application of UDI to SPP producers. From a practical perspective, the EU Commission guidance recommends that a SPP producer applies for registration on EUDAMED and obtains a Single Registration Number (SRN). Further guidance is available on the EU Commission [website](#).

Please note that a distributor will take on the responsibilities of the legal manufacturer and not the role of a SPP producer when undertaking any of the following activities while combining devices with other devices or products:

- Placing the combined items on the market in such a way that it not compatible with their original intended purpose;
- Placing the distributor registered name, trade mark and place of business in the place of the original OEMs;
- If including a non-CE marked component within the system or procedure pack;
- Sterilising in a manner that is not compatible with the original manufacturer's instructions;
- Placing a drug-device combination on the market where the drug is the principal mechanism of action (e.g. a co-packaged drug and syringe).

A distributor must then fulfil the obligations of a legal manufacturer, which include conducting the appropriate conformity assessments and registering with the HPRA. Where a distributor combines these products at the request of a health institution, Article 22 does not apply unless the health institution requests the distributor or SPP producer to undertake a specific activity that may change the intended purpose of the device such as sterilisation. In these instances, distributors or SPP producers should verify their obligations under Article 16.

The MDCG has also issued specific [guidance](#) on this topic and UDI requirements related to system and procedure packs.

8 PARTS AND COMPONENTS

Article 23 of the MDR outlines the obligations of economic operators intending to replace parts and components for devices that are defective or worn. An example of where this may apply is in the case of servicing and maintenance of medical devices in line with the manufacturer's guidelines. Repairs are permitted under MDR, however the requirements of Article 23 on parts and components must be fulfilled. This also extends to devices which have reached end of life, where the manufacturer is only providing replacement parts. Please note, where repairs are

conducted on single-use devices it may constitute reprocessing and such activities should only be carried out under the provisions allowed for by Member States.

Economic operators carrying out repairs as part of servicing and maintenance activities should do so in line with the manufacturer's guidelines and must prepare and keep documentation demonstrating there is no change to the safety, performance or intended purpose of the device. This documentation needs to be kept available for the competent authorities (i.e. the HPRA) and submitted upon their request. If your organisation is conducting repairs and the safety, performance or intended purpose of the device changes then the requirements of Article 16 of the MDR will apply and the organisation will assume the responsibilities of the legal manufacturer.

Typically, spare parts or components intended to repair existing components on a CE marked medical device are not considered medical devices in their own right. However, if these parts or components significantly change the performance or safety characteristics of the intended purpose of this device, these parts shall be considered devices in their own right. As a result, you must then treat them as you would any other device and fulfil your distributor obligations or checks in relation to these devices.

It is important to note the difference between operators carrying out repairs and those intending to fully refurbish medical devices. Repairing a medical device involves the replacement of original components or parts with new or reclaimed parts. As per Article 2 (31), 'fully refurbishing' is the complete rebuilding of a medical device already placed on the market or put into service, or the making of a new device from used medical devices. Under the MDR, fully refurbished medical devices will require CE marking and the entity carrying out the refurbishment will be considered as the 'manufacturer'. This manufacturer will be required to fulfil their obligations as laid out in the Regulations.

9 FIELD SAFETY CORRECTIVE ACTIONS (FSCA)

An FSCA is an action taken by a manufacturer to reduce a risk of death or serious deterioration in the state of health associated with the use of a medical device that is already placed on the market.

The following are examples of FSCAs:

- the return, withdrawal or recall of a medical device to the supplier or manufacturer (see also, section 9.1 for further details on medical device recalls)
- modification of the medical device
- modification to the clinical management of patients
- exchange of the medical device
- destruction of the medical device
- retrofit by the purchaser of the manufacturer's modification or design change

It is the manufacturer’s legal responsibility to inform Member States of an FSCA. An FSCA will always result in the issuance of a field safety notice (FSN) and in some instances will result in the issuance of an FSN and an additional HPRA safety notice (see figure 1 below).




Figure 1: Safety notice communication sequence

Nonetheless, a documented procedure for handling FSCAs notified by the manufacturer should be in place. In addition, the system established should be capable of handling the reporting of FSCAs where the distributor is responsible for initiating or reporting an FSCA. The document should clearly outline the various responsibilities with respect to the planning, conducting and reporting of the corrective action taken. The procedure should also clearly stipulate that the HPRA should be notified before any FSCA is undertaken and/or communicated onward.

Distributors should note that a device modification can include permanent or temporary changes to the labelling or instructions for use, software upgrades including those carried out by remote access and advice that changes the way the device is used (e.g. when an IVD manufacturer advises of a revised quality control procedure such as the use of third party controls or more frequent calibration or modification of control values for IVDs).

Further information regarding the vigilance system for medical devices is available on the [HPRA website](#).

9.1 Medical device recalls

A distributor should have a system in place to enable the swift and effective recall from the marketplace of defective and/or potentially harmful medical devices. In the event of a recall the responsibility of the distributor will depend on where in the supply chain they act for the medical device in question.

At a minimum, there should be an efficient and effective method for identifying customers supplied with a medical device subject to a recall. The recall procedure should be regularly challenged (recommended to be done at least once a year) to ensure that the process is effective and capable of tracing all customers and medical devices in a timely manner. This challenge may involve identifying a particular batch of a medical device and reconciling quantities received with those in stock and distributed to customers. A mock recall should be carried out annually if the company has participated in an actual recall during the previous year, which has used the same traceability system.

10 COMPLAINTS

In accordance with the Regulations, distributors who have received complaints or reports from healthcare professionals, patients or users about suspected incidents related to a device they have made available, must immediately forward this information to the manufacturer and, where applicable, their authorised representative and the importer. Distributors must keep a register of complaints. It is the responsibility of the distributor to keep the manufacturer (and, where applicable, the authorised representative and the importer) informed of any complaints or reports that may have an impact on product supplied to the market, and to provide them with any information upon their request.

A contract or technical agreement may be in place between the distributor and the manufacturer specifying responsibilities relating to the handling of complaints, incidents and vigilance.

The process to be followed upon receipt of a complaint, an event or incident report should be documented. A complaint, event and incident report register should be established.

All complaints, events or incidents received should be evaluated and categorised into either a quality, technical/service, vigilance or distribution related complaint depending on the nature of the report. Complaints specific to activities performed by the distributor should be investigated by the distributor. The decisions and actions taken as a result of a complaint, event or incident should be justified and recorded. Where complaints are not reported to the manufacturer or authorised representative/importer, this should be justified and documented. All communications regarding the complaint should be documented.

11 THE MANAGEMENT OF RETURNED MEDICAL DEVICES

A medical device should be considered to be a 'return' once it has left the premises of the supplying distributor and is subsequently returned to that premises. This may include the following examples:

- where a distributor supplies a customer with the incorrect medical device, which is subsequently returned
- where a customer returns a medical device to a distributor which they ordered in error
- where a medical device is received back to the premises of a distributor having never been received by the customer (e.g. because the customer's premises was closed)
- where a defective medical device is received back to the premises of a distributor for exchange or repair

Distributors should be extremely vigilant in their assessment of the suitability of returned medical devices to be placed back into saleable stock. Once the returned medical device has been placed back into saleable stock it may not be possible to distinguish between the returned medical device and the remainder of the stock even if the batch/lot number of the returned medical device was recorded. The distributor must be fully confident that the safety and

performance of the medical device has not been affected in any way whilst the medical device has been out of their care. Defective devices should be segregated from other stock and identified as defective.

When a return is received back it should be placed in a separate area so that there is no risk of it being returned to saleable stock prior to assessment in error. This separate area should be clearly segregated from saleable stock (either by physical means or by a validated computerised system).

All stages of the returns process should be documented. This documentation should allow all stages of the returns process to be traced, including the person conducting each stage/activity.

A suitably competent person should perform the checks on returned medical devices. If the medical device is to be rejected, then it should be placed into a reject area.

Personnel involved in the returns process should receive appropriate training and have sufficient experience in relation to the handling of medical devices to improve their ability to identify falsified medical devices.

The distributor must ensure that the correct storage conditions have been maintained during the period the medical device was outside of the distributor's control. There must be no reasonable possibility that the storage conditions were compromised during this period resulting in a risk to the performance and/or safety of the device.

There should be a register of returns in place which should include all medical device details and reasons for their return. The assessment performed on returned medical devices should be documented and should include the final disposition (i.e. approved for sale, quarantined, rejected or intended for destruction).

Special care must be exercised with the return of sterile medical devices and any medical devices requiring storage at low temperatures. Where a distributor decides to accept returns of medical devices requiring storage at low temperatures, the criteria for accepting these should be clearly documented and strictly adhered to in practice. The distributor must have documented evidence available for review confirming that the medical device was maintained within the cold chain for the entire time period during which it was outside of its control.

It is considered best practice if sterile medical device stock is only returned to saleable stock where there is no reasonable possibility that sterility has been compromised. If there is any doubt about their storage, then they must not be returned to saleable stock. Stocks of sterile products with broken seals, damaged packaging or where possible contamination is suspected must not be returned to saleable stock.

Distributors should be aware of the potential for falsified medical devices to enter the supply chain through the returns process. All relevant staff members should be made aware of this.

12 PERSONNEL AND STAFF TRAINING

Personnel involved in the distribution of medical devices should have the appropriate training and experience to adequately conduct their assigned duties. There should be sufficient competent personnel to carry out all the tasks for which the distributor is responsible.

It is suggested that each distribution centre appoint a person of appropriate training and experience who has primary responsibility for ensuring that regulatory responsibilities are met and that the quality system is effective. This person may also serve as the primary contact between the competent authority and the distributor.

Other duties that need to be carefully controlled and assigned to key personnel of appropriate training and experience, include the segregation, storage and release of returned stock to saleable stock and the handling of product recalls and field safety notices.

All training events should be documented in sufficient detail, including: a description of training, the duration and location of training, the training provider, who was trained and whether the trainee received the required level of competence with respect to the objective of the training event. For training on documents, for example procedures, the revision number of the document should be included.

Both trainer and trainee should sign off on all training records.

Training could include but is not limited to:

- defined responsibilities for personnel
- access to and training on documentation relevant to their role
- storage requirements
- labelling checks
- reporting of non-compliances
- operation of segregated areas to minimise the risk of mix-ups
- recall/withdrawal procedures
- complaints procedures
- regulatory requirements and related guidance documents

13 IMPLEMENTATION OF A QUALITY SYSTEM

In order to meet the legislative requirements and to ensure that only medical devices that comply with the legislation are made available for supply, it is suggested that distributors should have a quality system in place as a means of best practise. An effective quality system provides assurance that only medical devices which comply with legislative requirements are distributed, that non-compliant, defective or unsuitable medical devices can be detected, that traceability is

maintained and that non-conformances and the introduction of changes are controlled. There are various established standards for quality systems, with ISO 13485 being the most commonly used quality framework in the medical device industry. While adherence to such a framework can be beneficial in providing structure to a quality system, the quality system is not required to be officially accredited to any specific standard.

The quality system is the responsibility of the organisation's management. It requires their leadership and active participation, and should be supported by staff commitment. A person should be appointed by the management who should have clearly specified authority and responsibility for ensuring that a quality system is implemented and maintained. A quality manual or equivalent should be established in which all distribution activities should be clearly defined and systematically reviewed.

Standard operating procedures (SOPs) are recommended to clearly describe how key activities are carried out. In order to ensure that procedures are reflective of current requirements, a periodic review should be performed. This review should be documented and any recommendations should be implemented.

It is recommended that distribution of documents to staff is controlled in a manner such that only up-to-date and approved documents are available in relevant areas and obsolete copies should not be accessible. This may be achieved by maintaining a distribution list with records of procedures issued and retrieved, including the dates on which these activities took place. Superseded master copies of procedures should be maintained for a period of at least six years.

It is particularly important that SOPs relating to activities in certain areas (e.g. receipt of material at the goods inwards area) are available to staff in the relevant area for reference as required.

SOPs should describe the different operations which may affect the performance of the medical devices and could include but is not limited to the following:

- Training
- Documentation control
- Approval of suppliers and customers
- Order processing and deliveries
- Waste management
- Transport
- Audits
- Change control
- Complaints
- Quality risk management
- Management of Field Safety Corrective Actions
- Recalls and withdrawals
- Non-conformance management
- Corrective and preventive actions

At a minimum, distributors should keep a register of change controls, risk assessments, non-conformances, CAPAs, complaints, recalls and withdrawals and audits.

A distributor should ensure that the quality system in place complies with the legislative parameters governing the distribution of medical devices. In such cases the distributor should perform a gap analysis of the specific requirements relating to the distribution of medical devices against its current operational standard. Differences in approach or gaps identified should be addressed through the introduction of additional procedures or the amendment of existing procedures within the quality system as required.

13.1 Change control

The purpose of a change control system is to enable distributors to identify, document and assess changes that they introduce and how this may affect compliance and performance of the device. Such changes may include, for example: a change in the settings of a heating system or the relocation of a medical device storage area within a warehouse. Such changes may have a significant impact, and may potentially affect the performance of the medical device distributed. Therefore, it is vital that the change is conducted in a controlled manner.

It is recommended that a change control procedure and associated forms be implemented and appropriately documented. The purpose of the procedure should be to ensure that all changes to the operation are fully evaluated in terms of impact on the medical device's performance and traceability. The evaluation process should identify the areas impacted by the change, including processes, equipment, personnel, training, validation, quality system and regulatory implications. The required actions to give full effect to the change and ensure its implementation should be identified. In addition, changes should be formally approved **prior** to implementation. A system should be in place to ensure that all actions identified during the change control process have been completed within an appropriate timeline. The principles of quality risk management should be built into the change control process. Each change should be considered and a decision taken as to whether a risk assessment is required prior to approval of the change for implementation. The requirement for a risk assessment to be considered should be documented on the change control form.

13.2 Quality risk management

Quality risk management is a systematic process for the assessment, control, communication and review of risks to the safety and performance of the medical device. It is a valuable component of an effective quality system. Risk management may be used to assess the risk posed to the medical device as a result of a non-conformance from normal practices or to justify a proposed non-conformance from accepted practice. The use of risk management should be based on scientific knowledge, reason and practices. The level of detail contained within the risk management process should be reflective of the level of risk to the medical device. Implications for medical device safety and performance, security, traceability and follow up actions should be

detailed. Risk assessments should be carried out by competent personnel and should be reviewed and approved by the relevant responsible personnel. It is recommended that companies have a procedure in place and that training is provided to staff. All documentation for risk assessments performed should be available to an inspector during the course of an inspection.

ISO14971 'Medical devices – application of risk management to medical devices' is the most frequently used risk management standard within the medical device industry. While not all aspects will be relevant to distributors, there are concepts and methodologies that can be applied to medical device distributors. Other standards and risk management tools may be equally applicable.

13.3 Non-conformances, investigations and corrective and preventive actions (CAPA)

Non-conformances are deviations from the requirements of the Regulations and/or internal procedures. It is recommended that a procedure is in place outlining the process for identifying, documenting, investigating and closing non-conformances that occur and the timelines involved. An assessment should be performed to determine the medical device safety and performance implications and/or impact on the quality system.

It is important to note that distributors should commence the documentation of investigations immediately upon a potential non-conformance or incident being identified. Should the outcome of an investigation conclude that no non-conformance has occurred, then the documentation of the investigation should still be maintained and available to an inspector. An example of this may be where an investigation is commenced into stock discrepancies identified during stock counts where the outcome of the investigation is the location of the missing stock.

Investigations should aim to identify the root cause of the non-conformance. Corrective and preventive actions (CAPAs) may arise as a result of non-conformances, internal audits, observations or from other incidents.

A register of non-conformances should be maintained and all investigations, root cause identifications and resulting CAPAs documented. CAPAs should be subjected to regular review to ensure their full implementation and they should be subject to formal checks of their effectiveness.

The principles of quality risk management should be built into the non-conformance process. Each non-conformance should be considered and a decision taken as to whether a risk assessment is required. The requirement for a risk assessment to be considered should be documented on the non-conformance form.

13.4 Validation and qualification

To ensure the robustness of processes and equipment used in the distribution of medical devices it is best practice to validate or qualify these systems. Processes should be validated where the resulting output cannot be or is not verified by subsequent monitoring or measurement. Validation must demonstrate the ability of these processes to achieve planned results consistently. All validation exercises should be documented and the scope of the validation justified. Validation should be completed prior to the use of the system or process in the routine distribution and sale of the medical device (prospective validation). In exceptional circumstances, where this is not possible, it may be necessary to validate processes during routine distribution (concurrent validation).

Validation of critical processes

In order to validate a process, the distributor should first clearly describe/map the process (including the use of diagrams/flowcharts where relevant). The distributor should identify the critical steps within the process. Such steps may include goods-in, storage, order processing and dispatch. The distributor may further sub-divide these steps into the critical sub-steps involved in each individual process. For example, goods-in may be sub-divided into receipt on-site, delivery checking and acceptance, placing of goods into the goods-in area, detailed checking of medical devices, medical device acceptance or rejection and finally locating of goods to storage/picking/rejection area.

Once the distributor has identified the critical steps, a decision should be taken as to which steps require validation (or indeed whether all steps require validation). A risk management approach should be applied. Validation of processes should include an entire run through of a system from start to completion to ensure that each individual process does not have a negative impact on the following process. Records of the results and conclusion of validation and necessary actions arising from validation studies should be maintained.

The need for re-validation following a change to a system should be considered. Processes in use for some time should also be re-validated at a defined, justifiable period to demonstrate that they remain in a validated state.

Computer system validation

Before a computerised system (relating to distribution processes) is brought into use, it should be demonstrated, through appropriate validation or verification studies, that the system is capable of achieving the desired results accurately, consistently and reproducibly. The level of validation required will depend on the complexity of the system, whether it is a bespoke or 'off-the-shelf' system, and also the level of customisation performed on the system. The distributor should examine its systems and decide on the level of validation required using a risk management approach. There should be documentation available describing the computerised systems in use and the level of validation performed or planned to be performed. Software applications should be validated prior to initial use and, as appropriate, after changes.

Equipment qualification

In order to qualify a piece of equipment, a written protocol should be generated outlining how the qualification will be conducted. The protocol should describe the piece of equipment along with its critical functions and attributes. The protocol should describe how the correct operation of these critical functions and attributes will be verified along with acceptance criteria. The protocol should be reviewed and approved by the company.

Following completion of the qualification exercise, a report which cross-references the qualification and/or validation protocol should be prepared. The report should summarise the results obtained, comment on any non-conformances observed, and draw the necessary conclusions, including recommending changes necessary to correct deficiencies.

13.5 Outsourced activities

Technical or contractual agreements should be in place for all outsourced activities relating to distribution of medical devices. The technical agreements should at least describe the roles and responsibilities of both parties including details as applicable on transportation arrangements, receipt of goods, documentation, recalls, returns, customer complaints, suspected falsified medical devices, and management of non-conformances and changes. The technical agreement serves as a basis for defining the division of activities and responsibilities between the parties. However, it is important to highlight that the distributor retains ultimate responsibility for ensuring that the operations conducted are compliant with legal requirements. Technical agreements and procedures covering delegated activities may be reviewed during the course of a HPRA inspection.

13.6 Internal audits

In order to ensure and determine the effectiveness and compliance of the business with its quality system and legislation, it is recommended an internal audit or self-inspection programme is established. To this end, a risk based internal audit plan should be prepared that covers all the unit operations carried out in the course of business over a defined period of time. Any plan should be realistic and capable of achieving its objectives. Internal audits may cover individual unit operations (and be conducted more frequently) or numerous operations within their scope.

An internal audit or self-inspection log should also be established to record the scope of any internal audits carried out, the outcomes of any internal audit carried out and the actions (corrective and preventative) arising from any internal audit carried out. An audit log capturing such detail should prove useful with respect to trending the effectiveness of the quality system and determining the content of future self-inspection plans.

For all internal audits or self-inspections carried out, a detailed report should be prepared and any non-conformances identified should be closed out in a timely manner.

14 FURTHER INFORMATION

For queries on the distribution of medical devices in Ireland, contact the Health Products Regulatory Authority at the following address:

Compliance Department
Health Products Regulatory Authority
Kevin O'Malley House
Earlsfort Centre
Earlsfort Terrace
Dublin 2
D02 XP77

Telephone: +353-1-6764971

Fax: +353-1-6764061

Email: compliance@hpra.ie

Information is also available on the HPRA website at www.hpra.ie.

APPENDIX 1 FURTHER GUIDANCE

Selection of Articles from the Regulations

Medical Device Regulation

Article 10(8)

Manufacturers shall keep the technical documentation, the EU declaration of conformity and, if applicable, a copy of any relevant certificate, including any amendments and supplements, issued in accordance with Article 56, available for the competent authorities for a period of at least ten years after the last device covered by the EU declaration of conformity has been placed on the market. In the case of implantable devices, the period shall be at least 15 years after the last device has been placed on the market.

Upon request by a competent authority, the manufacturer shall, as indicated therein, provide that technical documentation in its entirety or a summary thereof.

A manufacturer with a registered place of business outside the Union shall, in order to allow its authorised representative to fulfil the tasks mentioned in Article 11(3), ensure that the authorised representative has the necessary documentation permanently available.

Article 10(11)

Manufacturers shall ensure that the device is accompanied by the information set out in Section 23 of Annex I in an official Union language(s) determined by the Member State in which the device is made available to the user or patient. The particulars on the label shall be indelible, easily legible, clearly comprehensible to the intended user or patient.

Article 13(3)

Importers shall indicate on the device or on its packaging or in a document accompanying the device their name, registered trade name or registered trade mark, their registered place of business and the address at which they can be contacted, so that their location can be established. They shall ensure that any additional label does not obscure any information on the label provided by the manufacturer.

IVD Regulation

Article 10(7)

Manufacturers shall keep the technical documentation, the EU declaration of conformity and, if applicable, a copy of the relevant certificate, including any amendments and supplements, issued in accordance with Article 51, available for the competent authorities for a period of at least ten years after the last device covered by the EU declaration of conformity has been placed on the market.

Upon request by a competent authority, the manufacturer shall, as indicated therein, provide that technical documentation in its entirety or a summary thereof.

A manufacturer with a registered place of business outside the Union shall, in order to allow its authorised representative to fulfil the tasks mentioned in Article 11(3), ensure that the authorised representative has the necessary documentation permanently available.

Article 10(10)

Manufacturers shall ensure that the device is accompanied by the information set out in Section 20 of Annex I in an official Union language(s) determined by the Member State in which the device is made available to the user or patient. The particulars on the label shall be indelible, easily legible and clearly comprehensible to the intended user or patient.

The information supplied in accordance with Section 20 of Annex I with devices for self-testing or near-patient-testing shall be easily understandable and provided in the official Union language(s) determined by the Member State in which the device is made available to the user or patient.

Article 13(3)

Importers shall indicate on the device or on its packaging or in a document accompanying the device their name, registered trade name or registered trade mark, their registered place of business and the address at which they can be contacted, so that their location can be established. They shall ensure that any additional label does not obscure any information on the label provided by the manufacturer.